

**CERTIFICATE OF CONFORMANCE
EUROPEAN COMMUNITY
COUNCIL DIRECTIVE 2014/30/EU**

Date of Issue: April, 2019

Issued By: Retlif Testing Laboratories
101 New Boston Road
Goffstown, NH 03045

Issued To: United Electronic Industries, Inc.
27 Renmar Ave.
Walpole, MA 02081

Reference: Retlif Report Number R-6246N-1

Retlif Testing Laboratories hereby acknowledges that compliance testing in accordance with the below listed standards was performed on a representative sample of the equipment listed below. Retlif Testing Laboratories further acknowledges that the test sample listed below was found to be in compliance with these standards.

This certificate is hereby issued to the above named grantee and is valid only for the equipment identified below.

Manufacturer: United Electronic Industries, Inc.
27 Renmar Ave.
Walpole, MA 02081

Equipment Tested: DNR-12-1G RACKtangle Ethernet I/O Rack

System Model Number: DNR-12-1G

System Serial Number: 0178191

Brand Name: United Electronic Industries

Product Type: Measurement, Control Equipment and Laboratory Use

Note(s): 1) See attached report R-6246N-1 for details and/or conditions pertaining to this certificate.
2) Conforms to the emissions requirements of EN 61326-1:2013; Clause 7.2:
CISPR 11: 2009, A1: 2010 Conducted Emissions, Group 1, Class A
CISPR 11: 2009, A1: 2010 Radiated Emissions, Group 1, Class A
3) Conforms to the immunity requirements of EN 61326-1:2013; Table 2:
IEC 61000-4-2: 2008 Electrostatic Discharge
IEC 61000-4-3: 2006, A1: 2007, A2: 2010 Radiated Immunity
IEC 61000-4-4: 2004, Cor2:2007, A1: 2010 EFT/Burst, Power Ports and I/O Ports
IEC 61000-4-5: 2005 Surge Immunity, Power Ports
IEC 61000-4-6: 2008 Conducted Immunity, Power Ports and I/O Ports
IEC 61000-4-11: 2004 Voltage Dips and Interrupts

**CERTIFICATE OF CONFORMANCE
EUROPEAN COMMUNITY
COUNCIL DIRECTIVE 2014/30/EU**

Date of Issue: March, 2020

Issued By: Retlif Testing Laboratories
101 New Boston Road
Goffstown, NH 03045

Issued To: United Electronic Industries, Inc.
27 Renmar Avenue
Walpole, MA 02081

Reference: Retlif Report Number R-6247N-1

Retlif Testing Laboratories hereby acknowledges that compliance testing in accordance with the below listed standards was performed on a representative sample of the equipment listed below. Retlif Testing Laboratories further acknowledges that the test sample listed below was found to be in compliance with these standards.

This certificate is hereby issued to the above named grantee and is valid only for the equipment identified below.

Manufacturer: United Electronic Industries, Inc.
27 Renmar Avenue
Walpole, MA 02081

Equipment Tested: DNF-4-1G FLATRACK Ethernet IO DAQ Platform

Model Number: DNF-4-1G

Serial Number: 223347

Brand Name: UEI

Product Type: Measurement, Control Equipment and Laboratory Use

Note(s): 1) See attached report R-6247N-1 for details and/or conditions pertaining to this certificate.
2) Conforms to the emissions requirements of EN 61326-1:2013; Clause 7.2:
CISPR 11: 2009, A1: 2010 Conducted Emissions, Group 1, Class A
CISPR 11: 2009, A1: 2010 Radiated Emissions, Group 1, Class A
3) Conforms to the immunity requirements of EN 61326-1:2013; Table 2:
IEC 61000-4-2: 2008 Electrostatic Discharge
IEC 61000-4-3: 2006, A1: 2007, A2: 2010 Radiated Immunity
IEC 61000-4-4: 2004, Cor2:2007, A1: 2010 EFT/Burst, Power Ports and I/O Ports
IEC 61000-4-5: 2005, Corr.: October 2009 Surge Immunity, Power Ports
IEC 61000-4-6: 2008 Conducted Immunity, Power Ports and I/O Ports
IEC 61000-4-11: 2004 Voltage Dips and Interrupts

**CERTIFICATE OF CONFORMANCE
EUROPEAN COMMUNITY
COUNCIL DIRECTIVE 2014/30/EU**

Date of Issue: February, 2020

Issued By: Retlif Testing Laboratories
101 New Boston Road
Goffstown, NH 03045

Issued To: United Electronic Industries, Inc.
27 Renmar Avenue
Walpole, MA 02081

Reference: Retlif Report Number R-6249N-1

Retlif Testing Laboratories hereby acknowledges that compliance testing in accordance with the below listed standards was performed on a representative sample of the equipment listed below. Retlif Testing Laboratories further acknowledges that the test sample listed below was found to be in compliance with these standards.

This certificate is hereby issued to the above named grantee and is valid only for the equipment identified below.

Manufacturer: United Electronic Industries, Inc.
27 Renmar Avenue
Walpole, MA 02081

Equipment Tested: DNA-PPC8-1G PowerDNA Gigabit Ethernet I/O Cube

Model Number: DNA-PPC8-1G

Serial Number: 162637

Brand Name: UEI

Product Type: Measurement, Control Equipment and Laboratory Use

Note(s): 1) See attached report R-6249N-1 for details and/or conditions pertaining to this certificate.
2) Conforms to the emissions requirements of EN 61326-1:2013; Clause 7.2:
CISPR 11: 2009, A1: 2010 Conducted Emissions, Group 1, Class A
CISPR 11: 2009, A1: 2010 Radiated Emissions, Group 1, Class A
3) Conforms to the immunity requirements of EN 61326-1:2013; Table 2:
IEC 61000-4-2: 2008 Electrostatic Discharge
IEC 61000-4-3: 2006, A1: 2007, A2: 2010 Radiated Immunity
IEC 61000-4-4: 2004, Cor2:2007, A1: 2010 EFT/Burst, Power Ports and I/O Ports
IEC 61000-4-5: 2005, +Corr. October: 2009 Surge Immunity, Power Ports
IEC 61000-4-6: 2008 Conducted Immunity, Power Ports and I/O Ports
IEC 61000-4-11: 2004 Voltage Dips and Interrupts

**CERTIFICATE OF CONFORMANCE
EUROPEAN COMMUNITY
COUNCIL DIRECTIVE 2014/30/EU**

Date of Issue: March, 2020

Issued By: Retlif Testing Laboratories
101 New Boston Road
Goffstown, NH 03045

Issued To: United Electronic Industries, Inc.
27 Renmar Avenue
Walpole, MA 02081

Reference: Retlif Report Number R-6250N-1

Retlif Testing Laboratories hereby acknowledges that compliance testing in accordance with the below listed standards was performed on a representative sample of the equipment listed below. Retlif Testing Laboratories further acknowledges that the test sample listed below was found to be in compliance with these standards.

This certificate is hereby issued to the above named grantee and is valid only for the equipment identified below.

Manufacturer: United Electronic Industries, Inc.
27 Renmar Avenue
Walpole, MA 02081

Equipment Tested: DNR-MIL 12-Slot Military Grade I/O Rack

Model Number: DNR-MIL-12-1G

Serial Number: 202219

Brand Name: UEI

Product Type: Measurement, Control Equipment and Laboratory Use

Note(s): 1) See attached report R-6250N-1 for details and/or conditions pertaining to this certificate.
2) Conforms to the emissions requirements of EN 61326-1:2013; Clause 7.2:
CISPR 11: 2009, A1: 2010 Conducted Emissions, Group 1, Class A
CISPR 11: 2009, A1: 2010 Radiated Emissions, Group 1, Class A
3) Conforms to the immunity requirements of EN 61326-1:2013; Table 2:
IEC 61000-4-2: 2008 Electrostatic Discharge
IEC 61000-4-3: 2006, A1: 2007, A2: 2010 Radiated Immunity
IEC 61000-4-4: 2004, Cor2:2007, A1: 2010 EFT/Burst, Power Ports and I/O Ports
IEC 61000-4-5: 2005, Corr.: October 2009 Surge Immunity, Power Ports
IEC 61000-4-6: 2008 Conducted Immunity, Power Ports and I/O Ports
IEC 61000-4-11: 2004 Voltage Dips and Interrupts

**CERTIFICATE OF CONFORMANCE
EUROPEAN COMMUNITY
COUNCIL DIRECTIVE 2014/30/EU**

Date of Issue: February, 2020

Issued By: Retlif Testing Laboratories
101 New Boston Road
Goffstown, NH 03045

Issued To: United Electronic Industries, Inc.
27 Renmar Avenue
Walpole, MA 02081

Reference: Retlif Report Number R-6360N-1

Retlif Testing Laboratories hereby acknowledges that compliance testing in accordance with the below listed standards was performed on a representative sample of the equipment listed below. Retlif Testing Laboratories further acknowledges that the test sample listed below was found to be in compliance with these standards.

This certificate is hereby issued to the above named grantee and is valid only for the equipment identified below.

Manufacturer: United Electronic Industries, Inc.
27 Renmar Avenue
Walpole, MA 02081

Equipment Tested: DNA-eCAT-800 EtherCAT based I/O Cube

Model Number: DNA-eCAT-800

Serial Number: 0223869

Brand Name: UEI

Product Type: Measurement, Control Equipment and Laboratory Use

Note(s): 1) See attached report R-6360N-1 for details and/or conditions pertaining to this certificate.
2) Conforms to the emissions requirements of EN 61326-1:2013; Clause 7.2:
CISPR 11: 2009, A1: 2010 Conducted Emissions, Group 1, Class A
CISPR 11: 2009, A1: 2010 Radiated Emissions, Group 1, Class A
3) Conforms to the immunity requirements of EN 61326-1:2013; Table 1:
IEC 61000-4-2: 2008 Electrostatic Discharge
IEC 61000-4-3: 2006, A1: 2007, A2: 2010 Radiated Immunity
IEC 61000-4-4: 2004, Cor2:2007, A1: 2010 EFT/Burst, Power Ports and I/O Ports
IEC 61000-4-5: 2005, +Corr. October: 2009 Surge Immunity, Power Ports
IEC 61000-4-6: 2008 Conducted Immunity, Power Ports and I/O Ports
IEC 61000-4-11: 2004 Voltage Dips and Interrupts

